

FEB 09 2012

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130172</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Freddie Wells</u>	ID# <u>4246</u> Date <u>1-13-12</u>
A Agency <u>Alaska State Troopers</u> Phone # <u>907-863-5111</u>	
Instrument Location <u>TOK, Alaska</u>	
B Alco S/N <u>X173017</u> Target Value <u>.083</u> High Pressure <u>600</u>	
Alco Test Values <u>.084</u> <u>.084</u>	
1st Alco 2nd Alco	
Signature <u>[Signature]</u> <u>COB</u>	
2/9/12	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5 day of April, 2012.

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130172

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

JANUARY 13, 2012

OPERATOR'S NAME:

WELLS/FREDDIE/L

OPERATOR'S NUMBER: 4246

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/NI :

VOC

O.L. #: 000000100

DEPT/AGENCY: TOKO

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .083

ALCO S/N: X173017

— BREATH ANALYSIS —

.083 ADJUSTED FOR 28.87 in		
ALCO TARGET	.080	13:01
BLANK TEST	.000	13:02
INTERNAL STANDARD	VERIFIED	13:03
ALCO TO 28.87 in	.084	13:03
BLANK TEST	.000	13:04
SUBJECT SAMPLE	.000	13:04
BLANK TEST	.000	13:05
ALCO TO 28.87 in	.084	13:05
BLANK TEST	.000	13:06

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

JANUARY 13, 2012

TIME 13:07

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	28.87 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ\]^_`abcdefghijklmnopqrstuvwxyz{|~
pgstuvwxyz{|~